Via Michelangelo Buonarroti, 52/B 20093 Cologno Monzese (Mi) Tel. (02) 253.80.88 r.a. - Telefax (02) 27300644 Sito Internet: www.cablas.com E-mail: cablas@cablas.com

## SCHERMATURE ANTIX - ACCESSORI PER RADIOLOGIA

## CE CONFORMITY DECLARATION

The undersigned "CABLAS S.r.l." office at Via Michelangelo Buonarroti 52/B 20093 Cologno Monzese (MI) - ITALY

DECLARES
Under own responsibility that the product

## **ANTIX SLIDING PANELS**

## Codes

P-PA-01 – Antix shielding with Pb mm 1	
P-PA-02 – Antix shielding with Pb mm 2	

To which this declaration relates is in conformity with the rules of good technique according to art. 72 D.L. 230 of 1995 which transposed the Directive EURATOM 80/836, 84/466, 89/618, 90/641, 92/3 in the field of ionizing radiation.

The lead is used for the initial smelting purposes only 99.9% in compliance with UNI 3165.

The thickness of the various protections, established in accordance with the UNI above, corresponds to what was specifically requested by the customer.

Cologno Monzese, year 2013

CABLAS S.R.L.
Via Michelangelo, 52/B 20093 COLOGIO MONZESE (MI)
S.M.: Via Somani 34 - COLOGIO M.S.E Tell R.A. 02 2538088 - Fax 02 27300644 Cod. Fisc. e P. IVA 11 9 0 3 2 3 0 1 5 6